

CERTYFIKAT ANALIZY



Komplet - wialki typ ND9 oranż z polem do opisu + nakrętki z septą biały Silikon/czerwony PTFE z nacięciem Alfatec™ Ultra Clean

Nr katalogowy producenta: SV022A222CSS

Nr katalogowy Alfachem: ALAA222CSS- KIT

ISO 9001 Quality Standard		
The product was manufactured under Membrane Solutions Quality Management System and is approved by an accredited registering body to the ISO 9001 Quality Systems Standard.		
Appearance of Specification		
Dimensional Test Items	Standards	Results
Volume	2	Pass
Vial OD	11.6±0.2mm	Pass
Mouth ID	6.25±0.2mm	Pass
Vial Height	32±0.5mm	Pass
Screw OD	9.5±0.2mm	Pass
Bottom Thickness	≥0.50mm	Pass
Chemical and Physical Performance		
Characteristics	Standards	Results
Content of B ₂ O ₃	≥8%g/g	Pass
Coefficient of Thermal Expansion	≤5×10 ⁻⁶ K ⁻¹	Pass
Hydrolytic Resistance of Glass Grains at 121°C	1	Pass
Heavy metals and Arsenic release	Content of Heavy metals and Arsenic is considerably below the limit values of European Ph. and China Ph.	

We hereby certify that the above stated sample vials are in compliance with all applicable instructions.

Osoba odpowiedzialna: **Małgorzata Rutkiewicz**

Dokument generowany elektronicznie i jest ważny bez podpisu i pieczętki



Alfachem Sp. z o.o. ul. Unii Lubelskiej 3, 61-249 Poznań, tel./fax: 61 875 30 66, tel. kom.: 516 840 149
biuro@alfachem.pl malgorzata@alfachem.pl

Copyright by AlfaChem 2013

CERTYFIKAT ANALIZY



Data produkcji: 23.05.2020

Data ważności: 23.05.2025

Osoba odpowiedzialna: **Małgorzata Rutkiewicz**

Dokument generowany elektronicznie i jest ważny bez podpisu i pieczętki



Alfachem Sp. z o.o. ul. Unii Lubelskiej 3, 61-249 Poznań, tel./fax: 61 875 30 66, tel. kom.: 516 840 149
biuro@alfachem.pl malgorzata@alfachem.pl

Copyright by AlfaChem 2013